

Analytical Test Report

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

Ablis Cran Blood Orange Kegs FD

Lab ID: B240457-02

Ablis CBD

METRC Batch ID: keg.cbo METRC Sample Tag: KEG.CBO Date Sampled: 04/26/24 Date Printed: 05/6/24

Potency Analysis

Analytical Method: JA-Potency-Proprietary BBLB SOP 009 - Potency Analysis by HPLC-UV-Vis

Cannabinoids	mg/g	LOQ Notes
THCA	< LOQ	0.002648
delta 9-THC	< LOQ	0.002648
delta 8-THC	< LOQ	0.002648
CBDA	< LOQ	0.002648
CBGA	< LOQ	0.002648
CBD	0.1432	0.002648
CBG	< LOQ	0.002648
CBN	< LOQ	0.002648
CBC	< LOQ	0.002648
THCV	< LOQ	0.002648
Total CBG	< LOQ	0.001324
Total Cannabinoids	0.1432	0.001324


Total THC < LOQ mg/g

Total CBD 0.1432 mg/g

<LOQ - Results below the Limit of Quantitation

ORELAP accredited cannabinoid analytes include only CBDA, CBD, THCA, delta-9-THC, and delta-8-THC. Acid form of THC/CBD are decarboxylated by heat, lose 12% of original mass as CO2. Result = *bioactive*

[&]quot;Total" Cannabinoid accounts for decarboxylation and moisture content. Total THC = [(THCA×0.877) + Δ 9THC] / (100%-MC)



Analytical Test Report

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

Ablis Cran Blood Orange Kegs FD

Ablis CBD

Laboratory ID: B240457-02

Quality Control Potency

Batch: B24D285 - BB Potency

Blank(B24D28	5-BLK1)						
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
delta 9-THC	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBGA	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBDA	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBD	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBN	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBG	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
delta 8-THC	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
CBC	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
Total CBG	< LOQ	0.01971	mg/g		04/29/24 09:33	04/29/24 10:28	
THCV	< LOQ	0.03943	mg/g		04/29/24 09:33	04/29/24 10:28	
LCS(B24D285-	BS1)						
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	100	0.03878	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
delta 9-THC	99.6	0.03878	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
CBDA	103	0.03878	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
CBD	102	0.03878	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
delta 8-THC	92.5	0.03878	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
LCS Dup(B24D	285-BSD1)						
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	99.1	0.03774	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
delta 9-THC	99.7	0.03774	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	
CBDA	98.9	0.03774	mg/g	90-110	04/29/24 09:33	04/29/24 10:28	

90-110

90-110

04/29/24 09:33

04/29/24 09:33

04/29/24 10:28

04/29/24 10:28

CBD

delta 8-THC

0.03774


0.03774

mg/g

mg/g

103

95.7



Analytical Test Report

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

Ablis Cran Blood Orange Kegs FD

Ablis CBD

Laboratory ID: B240457-02

Notes and Definitions

- B Analyte detected in method blank, but not associated samples.
- B2 Analyte detected in sample and associate methold blank.
- C Interference due to co-elution.
- D Initial result exceeded calibration range, reported data are based on analysis of a dilution.
- H Non-homogenous samle matrix affecting RPD and/or QC.
- I Manual Integration was performed.
- L Duplicate sample relative precent difference (RPD) exceeds QC limits.
- M Anomalous results due to matrix interference
- P Peaks manually split.
- Q1 QC out of limits but still oK
- Q2 Quality Control outside QC limits. Data considered estimate.
- Q3 CCV was above the acceptance criteria. Non-detect samples are considered acceptable.
- Q4 CCV was below the acceptance criteria, however the sample still exceeds the regulatory limit.
- R Marginal Exceedence.
- U Reported result is an estimate. The analyte was detected above the calibration range.
- X Problems with initial analysis, reported data are from reinjection of prepared sample.
- <LOQ Results below the Limit of Quantitation Compound not detected